Augmented Reality for Cryoablation Procedures

Hugo Talbot, Frederick Roy, Stéphane Cotin

► To cite this version:

Hugo Talbot, Frederick Roy, Stéphane Cotin. Augmented Reality for Cryoablation Procedures. SIG-GRAPH 2015, Aug 2015, Los Angeles, United States. hal-01180848

HAL Id: hal-01180848

<https://hal.inria.fr/hal-01180848>

Submitted on 31 Jul 2015

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Augmented Reality for Cryoablation Procedures

Hugo Talbot*, Frederick Roy and Stéphane Cotin
INRIA

Figure 1: Our physics-based simulation aims at estimating accurately the iceball and the isotherms during a cryoablation procedure

1 Introduction

Cryotherapy is a rapidly growing minimally invasive technique for the treatment of different kinds of tumors, such as breast cancer, renal and prostate cancer. Several hollow needles are percutaneously inserted in the target area under image guidance and a gas (usually argon) is then decompressed inside the needles. Based on the Thompson-Joule principle, the temperature drops down and a ball of ice crystals forms around the tip of each needle. Radiologists rely on the geometry of this iceball (273 K), visible on computer tomographic (CT) or magnetic resonance (MR) images, to assess the status of the ablation. However, cellular death only occurs when the temperature falls below 233 K. The complexity of the procedure therefore resides in planning the optimal number, position and orientation of the needles required to treat the tumor, while avoiding any damage to the surrounding healthy tissues.

This planning is currently done qualitatively, based on experience, and can take several hours, with a result that is often different from the expected one. To solve this important limitation of cryotherapy, a few planning systems have been proposed in the literature. Currently, commercial systems are nearly non-existent, and emerging tools are limited to a visualization of the isotherms obtained for each needle in ideal conditions (usually in a gel). They do not account for any influence of the soft tissue properties, the presence of blood vessels, or the combined effect of multiple needles highlighted in [Young et al. 2012]. As a consequence, large safety margins over 5 mm are defined, as detailed in [Georgiades et al. 2012].

To address this challenge, our method extracts information from medical images (CT or MR) and allows to assess different strategies with an augmented visualization of the resulting iceball and the associated isotherms, as illustrated in Fig. 1.

2 Our Approach

To estimate the heat transfer, our work relies on the Pennes' model [Pennes 1948], based on the enthalpy method. This physio-

logical approach accounts for blood perfusion in living tissue, key in the estimation of the iceball. Our heat transfer model is implemented using the Finite Element method. Starting from medical images (CT or MR), we define a region of interest around the tumor and create hexahedral elements from the image voxels. Blood vessels can easily be extracted from the images and incorporated within the simulation.

The diffusion term involves the integral of the derivative of the shape function ∇N_i , which is computationally-demanding. We propose a pre-computation of these derivatives using the regular feature of our hexahedra. The diffusion term can then be expressed as a sparse matrix and saved using a compressed row storage. Compared to linear tetrahedra, our approach reveals to be more accurate using the same number of elements. Implemented on GPU, the performance of our approach is currently being evaluated, and already shows promising results.

For each configuration of the needles, our method allows to compute the resulting iceball as well as the 233 K isotherm, using a marching cube algorithm. This work therefore provides the radiologists with a substantial visual guidance before the operation. Using our work, the clinician can thus test interactively various ablation strategies and evaluate the resulting 233 K isotherm. Moreover, our numerical simulation may allow to extend the use of the cryoablation technique to non-expert radiologists. The Fig. 1 shows the estimation of the iceball (273 K isotherm) in white, but also the 233 K isotherm, ensuring cellular death, in red.

As future work, we intend to use this efficient GPU simulation in order to build an automated planning system. Using the predictive power of our simulation, this system would estimate the optimal positions and orientations of each needle.

References

- GEORGIADES, C., RODRIGUEZ, R., AZENE, E., WEISS, C., CHAUX, A., AND *et al.* 2012. Determination of the nonlethal margin inside the visible ice-ball during percutaneous cryoablation of renal tissue. *Cardiovasc Intervent Radio*, pp. 1–8.
- PENNES, H. 1948. Analysis of tissue and arterial blood temperatures in the resting human forearm. *Journal of applied physiology* 1, 2, pp. 93–122.
- YOUNG, J. L., MCCORMICK, D. W., KOLLA, S. B., ET AL. 2012. Are multiple cryoprobe additive or synergistic in renal cryotherapy? *Urology* 79, 2, 484–e1.

*e-mail:hugo.talbot@inria.fr