

An efficient gradient-based method for differential-interference-contrast microscopy

Simone Rebegoldi, Lola Bautista, Marco Prato, Luca Zanni, Laure Blanc-Féraud, Arturo Plata

► To cite this version:

HAL Id: hal-01426337 https://hal.inria.fr/hal-01426337

Submitted on 4 Jan 2017

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

An efficient gradient-based method for differential-interference-contrast microscopy

Simone Rebegoldi, Lola Bautista, Marco Prato, Luca Zanni, Laure Blanc-Féraud, Arturo Plata

Università di Modena e Reggio Emilia, Via Campi 213/B, Modena, Italy Universidad Industrial de Santander, 680002 Bucaramanga, Colombia Université Nice Sophia Antipolis, 06903 Sophia Antipolis, France simone.rebegoldi@unimore.it, bautista@i3s.unice.fr, marco.prato@unimore.it, luca.zanni@unimore.it, laure.blanc_feraud@inria.fr, aplatag@yahoo.com

Keywords. DIC microscopy; inverse problems; gradient-descent method.

Differential-interference-contrast (DIC) microscopy is an optical microscopy technique widely used in biology to observe unstained transparent specimens, in which a two-dimensional image is formed from the interference of two waves that have a lateral differential displacement (shear) and are phase shifted relative one to each other. Following the rotational-diversity model proposed in [1], one is interested in retrieving the specimen's phase function from a set of DIC intensity images acquired at different rotations of the specimen. This highly nonlinear, illposed problem is solved by adopting a least squares approach and thus looking for a regularized solution of a smooth nonconvex optimization problem.

As already done in [1], one can address the DIC problem by means of a nonlinear conjugate gradient method, which is particularly suited for least squares problems. However, the computation of the line search parameter at each iteration may require several evaluations of both the function and its gradient in order to ensure convergence [2], which significantly increases computational time when such evaluations are time-consuming, as is the case of the DIC problem.

In this light we propose an efficient gradient-descent method for the estimation of a specimen's phase function from polychromatic DIC images. The method minimizes the sum of a nonlinear least-squares discrepancy measure and a smooth approximation of the total variation and exploits a recent updating rule for the choice of the step size [3]. Numerical simulations on two computer-generated objects show significant improvements in terms of efficiency and stability with respect to widely used conjugate gradient methods.

References

- [1] Preza C. (2000) Rotational-diversity phase estimation from differential interference contrast microscopy images. J. Opt. Soc. Am. A, Vol. 17, no. 3, pp. 415–424.
- [2] J. C. Gilbert, J. Nocedal. (1992) Global convergence properties of conjugate gradient methods for optimization. SIAM J. Optim., Vol. 2, No. 1, pp. 21–42.
- [3] R. Fletcher. (2012) A limited memory steepest descent method. Math. Program., Vol. 135, No. 1–2, pp. 413–436.