

Study on Plant Nutrition Indicator Using Leaf Spectral Transmittance for Nitrogen Detection

Juanxiu Hu, Dongxian He, Po Yang

► **To cite this version:**

Juanxiu Hu, Dongxian He, Po Yang. Study on Plant Nutrition Indicator Using Leaf Spectral Transmittance for Nitrogen Detection. Daoliang Li; Yande Liu; Yingyi Chen. 4th Conference on Computer and Computing Technologies in Agriculture (CCTA), Oct 2010, Nanchang, China. Springer, IFIP Advances in Information and Communication Technology, AICT-347 (Part IV), pp.504-513, 2011, Computer and Computing Technologies in Agriculture IV. <10.1007/978-3-642-18369-0_60>. <hal-01564825>

HAL Id: hal-01564825

<https://hal.inria.fr/hal-01564825>

Submitted on 19 Jul 2017

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Study on Plant Nutrition Indicator Using Leaf Spectral Transmittance for Nitrogen Detection

Juanxiu Hu¹, Dongxian He^{1,1}, and Po Yang²

¹ Key Lab. of Agriculture Engineering in Structure and Environment, College of Water Conservation and Civil Engineering, China Agricultural University, Beijing, P. R. China

² Beijing Lighting valley Technology Company, Beijing, P. R. China

Abstract. The low fertilizer utilization at growing season and environment pollution caused by unreasonable fertilization are becoming global outstanding problems in agricultural production. Scientific and reasonable fertilization based on rapid and nondestructive plant nutrient detection will be a valuable solution for solving above problems. In this study, spectral transmittance in wavelength ranged from 300 to 1100 nm, chlorophyll content and nitrogen content of rice and cucumber leaves treated with culture solution in five different nitrogen levels were measured. According to the correlation analysis between them, 560, 650, and 720 nm as feature wavelengths and 940 nm as reference wavelength were determined for nitrogen detection. Correlation analysis between 21 spectral feature parameters composed by the transmittance at above wavelength, the leaf chlorophyll content and nitrogen content, and combined with their regression examination indicated that spectral feature parameters of $(T_{940} - T_{560}) / (T_{940} + T_{560})$, $\log(T_{940}/T_{560})$ and $\log(T_{940}/T_{650})$ are useful to conduct plant nutrient diagnosis with less than 8% relative error in rice and cucumber leaves. Therefore, the above spectral feature parameters as plant nitrogen indicators can be used to estimate the chlorophyll content and nitrogen content, furthermore support for non-destructive plant nutrient detection and fertilizer recommendation based on testing soil.

Keywords: Chlorophyll content, Nitrogen content, Spectral feature parameter, Spectral transmittance

1 Introduction

Nitrogen is an essential nutrient factor for growth and quality of plants. The applying of nitrogen fertilizer has been growing persistently in order to increase the crop yield. However, the environmental pollutions caused by low seasonal utilization of nitrogen fertilizer and the unreasonable applications of chemical fertilizer have become outstanding issues in agricultural production [1, 2]. The amount of nitrogen in soil is closely related to the nitrogen nutrition status of plant. The scientific nitrogen fertilization based on evaluating nitrogen nutrition status of plant is an effective approach to solve the problem above.

¹ Corresponding email: he_dongxian@hotmail.com

Chlorophyll content and nitrogen content are important indications for evaluating nitrogen nutrition status of plant. Estimating chlorophyll content and nitrogen content by methods of chemical analysis though presenting reliable results, cannot meet the requirement of guiding fertilizer promptly because of the vary complicated and time-consuming process [3]. The application of remote sensing technology in agriculture has provided a new means for rapid and non-destructive diagnosis of plant nitrogen nutrition [4]. There are lots of studies focusing on diagnosis of plant nitrogen nutrition based on the spectral properties of plant were carried out. Spectral reflectance of plant leaves in the vicinity of 550 and 670 nm were highly relevant to chlorophyll and nitrogen contents by measuring the reflectance of leaves or canopy [5, 6]. Reflectance in 550 nm and 710 nm could be used to better estimate the nitrogen content in corn and $R(550\sim600) / R(800\sim900)$ could be sensitive to reflect the nitrogen stress in corn. The linear combinations of reflectance in 620 and 760 nm had good correlation with nitrogen contents in rice leaves [7].

As to studies on plant spectral transmittance and nitrogen nutrition, an in-vivo detection device was developed for estimation plant chlorophyll content based on the difference in leaf spectral transmittance between in 660-690 nm and in 760-1100 nm [8]. Spectral characteristic parameter $\log(I_{660-690}/I_{760-1100})$ was adopted in the device to estimate leaf chlorophyll content, in which I₆₆₀₋₆₉₀ and I₇₆₀₋₁₁₀₀ were the intensity of light within corresponding wavebands having transmitted through leaves. Minolta, a Japanese company, then improved the algorithm and developed a portable chlorophyll meter. The spectral feature parameter adopted in the meter was $\log(T_{940}/T_{650})$, namely, SPAD (Soil and Plant Analysis Development) value, in which T₉₄₀ and T₆₅₀ were the transmittance in corresponding wavelengths. The portable chlorophyll meter has been widely applied in rapid and non-destructive diagnosis of nitrogen nutrition in rice, maize, wheat, cotton and other field crops [9-12]. It also had a good performance in application for horticultural plants and xylophyta such as sweet potato, rape, peach, and maple, etc [13-16].

Many studies have been conducted on nitrogen nutrition diagnosis with the plant reflection spectrum, while the research on plant spectral transmittance is mostly based on the application of the portable chlorophyll meter. In this study, the feature band in direct relation with nitrogen nutrition of plants was determined based on transmittance spectra of rice and cucumber leaves cultivated in nutrient solution of different levels of nitrogen fertilizer. The correlation analysis and regression estimation error analysis were carried out of established spectral feature parameters and chlorophyll content as well as nitrogen contents, so as to determine plant nutrition indicators for nitrogen detection through estimating chlorophyll content and nitrogen content in leaves.

2 Materials and Methods

2.1 Cultivation of Plants

In this study, the cultivation of plants were performed in a environment controlled greenhouse using cucumber (*Cucumis sativus* L., cv., Zhongnong No.8 and Beijing 203) and rice (*Oryza sativa* L., cv., Wuyujing No. 3) as model plant during the period

from 2008 to 2009. Cucumber and rice seedlings were transplanted to 5 cultivation beds of the same dimension (250 cm × 60 cm × 40 cm), and 16 plants in each bed. A homogenous mixture of vermiculite, grass peat and perlite at a proportion of 3:1:1 was used as the cultivation base. The temperature and light intensity in greenhouse were optimally controlled through pad and fan cooling system and sunshade screen.




Fig. 1 The status of cucumber and rice cultivation.

Cucumber and rice plants were treated with culture solution in 5 different nitrogen levels. With the nitrogen level of nutrient solution formulation for cucumbers from Yamazaki Japan as 100%, content of nitrogen in the nutrient solutions were increased or decreased by 50% and 100% respectively while content equality of other major nutrient elements was ensured as far as possible (Table 1). Then the growing areas, according to the different nitrogen levels, were marked as N0, N50, N100, N150 and N200. The pH of culture solution was adjusted to 6.5-7.5. Pumps and timers were applied in automatic timing irrigation of the nutrient solution.

Table 1 Nutrient solution in different treatments for cucumber and rice cultivation

	The content of chemical in nutrient solutions (Unit: mg L ⁻¹)				
	N0	N50	N100	N150	N200
Ca(NO ₃) ₂ 4H ₂ O	0	0	826	826	826
CaCl ₂ 2H ₂ O	516	516	0	0	0
KNO ₃	0	607	607	607	607
KCl	370	0	0	0	0
KH ₂ PO ₄	136	0	0	0	0
NH ₄ H ₂ PO ₄	0	115	115	115	115
NH ₄ NO ₃	0	0	0	280	560
MgSO ₄ 7H ₂ O	483	483	483	483	483
	The content of major nutrient elements (Unit: mmol L ⁻¹)				
	N0	N50	N100	N150	N200
N	0	7	14	21	28
P	1	1	1	1	1
K	6	6	6	6	6
Ca	3.5	3.5	3.5	3.5	3.5
Mg	2	2	2	2	2

2.2 Measured Parameters and Measurement Methods

During the tillering stage of rice and fast growth period of cucumber, five samples of plant leaves in each treatment area were selected respectively for parameters measurement of the spectral property and nitrogen nutrition indices, including the spectral transmittance of leaf within the wavelength range of 300-1100 nm, and the chlorophyll content and nitrogen content of leaf.

(1) The determination of spectral transmittance

Spectral transmittance of rice or cucumber leaves was determined using an integrating sphere measurement device in the large sample chamber of spectrophotometer (UV3150, Shimadzu Co., Japan). The measurement wavelength range was 300-1100 nm with the interval of 1nm. The light source switch point was set at 360 nm and the raster switch point at 820 nm while the slit width was 20 nm. The measurement position was basically in both sides of the main vein and middle of the leaf. For each leaf, 4 different positions were selected for determination of spectral transmittance, so each leaf was measured for 4 times.

(2) The determination of chlorophyll content

0.05g rice leaves (0.20g cucumber leaves) were weighed, cut into filaments and put into test tubes. After adding 10mL mixture of 1:1 ethanol-acetone, the tubes were placed in refrigerator at 4°C for 16-18h in the dark for extracting chlorophyll of leaves. The extract was uniformly shaken and then transferred into a clean cell for determination of chlorophyll content. With the ethanol-acetone (1:1) mixture as the reference, absorbance values at 663 and 645nm of the extract were measured by a spectrometer (UV3150, Shimadzu Co., Japan). The chlorophyll content was calculated by the correction formula of Arnon method [17].

(3) The determination of nitrogen content

0.10g rice or cucumber leaves were weighed to a 100mL digestion tubes and wetted with a little water dripped in, and then 4mL of concentrated H_2SO_4 was added. The tubes were covered with small curved neck funnels to place overnight after shaken gently. The next day, the tubes were put into a digestion furnace to heat firstly at 180°C for an hour, and then at 300°C for 2 hours. When the solution became brown-red, the tubes were moved away and cooled for 3min. Then 15 drops of 30% H_2O_2 were added into the solution. The tube was put into the digestion furnace to heat at 300°C again for 15 min after shaken. Then 10 drops of 30% H_2O_2 was added after the tubes were moved away to cool for 3min again. The above procedures were repeated for times with the amount of 30% H_2O_2 gradually decreased until the solution became colorless or bright. Finally, the tubes were heated for 20min in order to remove residual H_2O_2 . After thoroughly cooled, the digestion solution was arranged on a Kjeldahl type automatic azotometer (KDY-9830, Beijing Tongrunyuan Electromechanical Technology Co., Ltd., China) for distillation titration to determine the nitrogen content.

2.3 Processing and Analysis of Data

The mean value of four measurements of each leaf sample's spectral transmittance within 300-1100 nm was adopted. The correlation analysis of the leaf spectral

transmittance, the chlorophyll content, and the nitrogen content was conducted with the SPSS software, so as to determine the feature wavelength and reference wavelength for nitrogen nutrition diagnosis in plants. With the correlation analysis of spectral feature parameters established by the spectral transmittance of characteristic wavelength and reference wavelength, the leaf chlorophyll content, and the leaf nitrogen contents, plant nitrogen nutrition indicators were determined and tested.

3 Results and Analysis

3.1 The Spectral Transmission of Plant Leaves




Fig. 2 Leaf spectral transmittances of cucumber and rice leaves treated with culture solution in different nitrogen levels.

Spectral transmittances of cucumber and rice leaves were changed basically with the same trend of the wavelengths in different nitrogen nutrition levels. In the range of photosynthetically active radiation wave band (400-700 nm), the leaf spectral transmittance formed wave crest at 550 nm and shaped up wave trough at 680 nm. At the same time, the leaf spectral transmittance reduced with the increasing nitrogen. Spectral transmittances of leaves were all between 40~50% in near-infrared wave band (800-1100 nm), in which the range of variation was little.

The spectral properties of plant leaves are mainly formed by absorption, reflection and transmission of light towards mesophyll cells, chlorophyll, moisture content and other biochemical components in the leaves. In the wave band of 400-700 nm, the spectral properties of leaves were mainly affected by chlorophyll. Due to strong absorption of red light and little absorption of green light by leaf chlorophyll, a wave crest was formed around 550 nm while a wave trough of low transmittances around 680 nm in leaf spectral transmittance. In fast growth period of plants, with the increasing of fertilizer nitrogen, the chlorophyll and nitrogen content also increased so that leaves could absorb more light at 400-700 nm for photosynthesis, therefore

spectral transmittance of plant leaves in this wave band decreased with the increasing of fertilizer nitrogen. In the near-infrared wave band (800-1100 nm), the spectral properties of leaves were mainly influenced by the tissue structure in leaves, while the impact of chlorophyll could be negligible. The spectral transmittance of leaf in the near-infrared wave band was 40-50%, which is in line with the results of plant leaf spectrum by other researchers referring to that the absorptivity of leaf was generally less than 10% and reflectivity was about 50% [18-19].

3.2 The Correlation between Spectral Transmittance of Leaves and Chlorophyll Contents as well as Nitrogen Contents




Fig. 3 Correlation coefficient between leaf spectral transmittance (LST), chlorophyll content (CC) and nitrogen content (NC).

The significant negative correlation between spectral transmittance in the waveband of 500-720 nm and leaf chlorophyll content as well as nitrogen content was found in rice and cucumber leaves with the correlation coefficients ranging from -0.8 to -0.9 ($P=0.01$). Within the waveband of 900-1100 nm, there was not a significant correlation between spectral transmittance and chlorophyll content as well as nitrogen content.

Integrated spectral feature parameters, such as difference value and ratio of spectral parameters between feature waveband and reference waveband were adopted in order to minimize the impact of structure and thickness of plant leaves on the spectral data analysis thereby enhancing the accuracy of estimations of chlorophyll and nitrogen content^[6, 7]. For instance, spectral feature parameter of $\log(T_{940}/T_{650})$ was used in the portable chlorophyll meter (SPAD-502) produced by Japanese company Minolta to estimate leaf chlorophyll content. A portable meter developed by Fuchigami et al for plant nutrient detection adopts T_{940}/T_{560} and T_{940}/T_{720} [20]. Based on the correlation between spectral transmittance and chlorophyll content as well as nitrogen content and relevant studies, 560, 650, 720 nm were selected as typical feature wavelengths in the 500-720 nm waveband, and at the same time 940 nm was chosen within the wave band of 900-1100 nm as the reference wavelength to establish 21 spectral feature parameters (Table 2) for nitrogen nutrition diagnosis of plants.

3.3 The Correlation between Spectral Feature Parameters of Leaves and Chlorophyll Contents as well as Nitrogen Contents

The spectral feature parameters the correlation coefficient between which and chlorophyll content in rice was over 0.94 (P=0.01) were as follows: $T940/T560$, $(T940 - T560) / (T940 + T560)$, $(1/T560 - 1/T940) / (1/T940)$, $(T940 - T560) / T940$, $\log (T940/T560)$, $T940/T650$, $(1/T650 - 1/T940) / (1/T940)$, $\log (T940/T650)$. Moreover, the correlation coefficient between all spectral feature parameters mentioned above and nitrogen contents in rice was over 0.91 (P=0.01) (Table 2). The spectral feature parameters the correlation coefficient between which and chlorophyll content in cucumber was over 0.91 (P=0.01) were as follows: $(T940 - T560) / (T940 + T560)$, $\log (T940/T560)$, $(T940 - T650) / (T940 + T650)$, $\log (T940/T650)$. The correlation coefficient between the two former spectral feature parameters and nitrogen content in cucumber was 0.87 and 0.88 respectively, while that of the latter two and nitrogen content was over 0.90 (P=0.01). It was indicated that the spectral feature parameters of $(T940 - T560) / (T940 + T560)$, $\log (T940/T560)$, and $\log (T940/T650)$ had a high correlation with chlorophyll and nitrogen contents in cucumber and rice.

Table 2 Correlation coefficient between spectral feature parameters and chlorophyll content as well as nitrogen content

	spectral feature parameter and chlorophyll content		spectral feature parameter and nitrogen content	
	rice	cucumber	rice	cucumber
$T940/T560$	0.94	0.90	0.93	0.87
$(T940 - T560) / (T940 + T560)$	0.95	0.91	0.92	0.87
$(T940 - T560) / T940$	0.94	0.90	0.91	0.87
$T940 - T560 - T940 \times T560$	0.92	0.89	0.88	0.86
$(1/T560 - 1/T940) / (1/T940)$	0.94	0.90	0.93	0.87
$\log (T940/T560)$	0.95	0.91	0.93	0.88
$\log (T940/T560) / \log (T940)$	-0.92	-0.88	-0.87	-0.84
$T940/T650$	0.94	0.89	0.92	0.87
$(T940 - T650) / (T940 + T650)$	0.91	0.91	0.89	0.90
$(T940 - T650) / T940$	0.90	0.90	0.88	0.89
$T940 - T650 - T940 \times T650$	0.76	0.78	0.71	0.76
$(1/T650 - 1/T940) / (1/T940)$	0.94	0.89	0.92	0.87
$\log (T940/T650)$	0.94	0.92	0.92	0.90
$\log (T940/T650) / \log (T940)$	-0.88	-0.89	-0.84	-0.88
$T940/T720$	0.90	0.86	0.88	0.81
$(T940 - T720) / (T940 + T720)$	0.90	0.86	0.88	0.81
$(T940 - T720) / T940$	0.90	0.86	0.88	0.80
$T940 - T720 - T940 \times T720$	0.89	0.85	0.87	0.80
$(1/T720 - 1/T940) / (1/T940)$	0.90	0.86	0.88	0.80
$\log (T940/T720)$	0.90	0.86	0.89	0.81
$\log (T940/T720) / \log (T940)$	-0.92	-0.84	-0.88	-0.78

3.4 The Experimental Examination of Spectral Feature Parameters

Linear regression analysis was carried out of spectral feature parameters, namely $(T940 - T560) / (T940 + T560)$, $\log (T940/T560)$ and $\log (T940/T650)$, and chlorophyll as well as nitrogen contents in rice and cucumber measured by chemical methods. Chlorophyll and nitrogen contents were estimated through the regression formula and compared with the values measured practically (Table 3). It was found that the relative errors of chlorophyll and nitrogen contents in rice estimated by $(T940 - T560) / (T940 + T560)$, $\log (T940/T560)$, and $\log (T940/T650)$ were below 5.4%. As regards to chlorophyll contents in cucumber, the relative errors were below 4.5%, and the relative error of its nitrogen content was less than 7.7%. These results indicated that the relative errors of estimated leaf chlorophyll and nitrogen contents with spectral feature parameters of $(T940 - T560) / (T940 + T560)$, $\log (T940/T560)$, and $\log (T940/T650)$, could meet the requirements of testing the plant nitrogen nutrition in vivo. Thus, they are suitable for use indicators of plant nitrogen nutrition. Among the spectral feature parameters mentioned above, $(T940 - T560) / (T940 + T560)$ was the same as was adopted in the plant nutrition detector developed by Fuchigami et al (2007) while $\log (T940/T650)$ was the same as used in the portable chlorophyll meter (SPAD-502) produced by Minolta Corporation.

Table 3 Relative error of estimated chlorophyll and nitrogen content using spectral feature parameter (n=30)

	Relative error of estimated chlorophyll content (%)		Relative error of estimated nitrogen content (%)	
	Rice	Cucumber	Rice	Cucumber
$(T940 - T560) / (T940 + T560)$	5.3±3.7	4.4±3.5	5.2±3.4	7.6±5.0
$\log (T940/T560)$	5.2±3.7	4.5±3.5	4.7±3.0	7.7±5.0
$\log (T940/T650)$	5.1±3.7	4.5±3.9	5.4±3.4	6.5±4.6

4 Conclusions

Spectral transmittance in wavelength ranged from 300 to 1100 nm, chlorophyll content and nitrogen content of rice and cucumber leaves treated with culture solution in five different nitrogen levels were measured. According to the correlation analysis between them, 560, 650, and 720 nm as feature wavelengths and 940 nm as reference wavelength were determined for nitrogen detection. Correlation analysis between 21 spectral feature parameters composed by the transmittance at above wavelength, the leaf chlorophyll content and leaf nitrogen content, and combined with their regression examination indicated that spectral feature parameters of $(T940 - T560) / (T940 + T560)$, $\log (T940/T560)$ and $\log (T940/T650)$ are useful to conduct the plant nutrient diagnosis with less than 8% relative error in rice and cucumber leaves. Therefore, the above spectral feature parameters as plant nitrogen indicators can be used to estimate the chlorophyll content and nitrogen content.

Acknowledgments. This work has been supported by National High Technology Research and Development Program of China (2006AA10A302) and (2006AA10Z206).

References

1. Zeng, X.B., Li, J.M.: Fertilizer Application and its Effect on Grain Production in Different Counties of China (in Chinese). *Scientia Agricultura Sinica* 37, 387--392 (2004)
2. Ju, X.T., Xing, G.X., Chen, X.P., Zhang, S.L., Zhang, L.J., Liu, X.J., Cui, Z.L.: Reducing Environmental Risk by Improving N Management in Intensive Chinese Agricultural Systems. *Proceedings of the National Academy of Science* 106, 3041--3046 (2009)
3. Jiao, W.J., Min, Q.W., Lin, K., Zhu, Q.K., Zhang, J.J.: Progress and Perspective on Nutrition Diagnosis of Plant Nitrogen (in Chinese). *Chinese Agricultural Science Bulletin* 22, 351--355 (2006)
4. Tian, Y.C., Zhu, Y., Yao, X., Liu, X.J., Cao, W.X.: Non-destructive Monitoring of Crop Nitrogen Nutrition Based on Spectral Information (in Chinese). *Chinese Journal of Ecology* 26, 1454--1463 (2007)
5. Thomas, J.R., Oerther, G.F.: Estimating Nitrogen Content of Sweet Pepper Leaves by Reflectance Measurements. *Agronomy Journal* 64, 11--13 (1972)
6. Filella, L., Serra, L., Penuelas, J.: Evaluating Wheat Nitrogen Status with Canopy Reflectance Indices and Discriminant Analysis. *Crop Science* 35, 1400--1405 (1995)
7. Shibayama, M., Akiyama, T.: A Spectroradiometer for Field Use VII. Radiometric Estimation of Nitrogen Levels in Field Rice Canopy. *Japanese Journal of Crop science* 55, 433--438 (1986)
8. Watanabe, S., Kusunuki, T.: Method of and Device for Measuring Chlorophyll of Living Leaves. Patent, Japan (1981)
9. Turner, F.T., Jund, M.F.: Chlorophyll Meter to Predict Nitrogen Topdress Requirement for Semidwarf Rice. *Agronomy Journal* 83, 926--928 (1991)
10. Wu, L.H., Tao, Q.N.: Nitrogen Fertilizer Application Based on the Diagnosis of Nitrogen Nutrition of Rice Plants (*Oryza sativa* L.) Using Chlorophyll Meter (in Chinese). *Journal of Zhejiang Agricultural University* 25, 135--138 (1999)
11. Piekielek, W.P., Fox, R.H.: Use of a Chlorophyll Meter to Predict Side-dress Nitrogen Requirements for Maize. *Agronomy Journal* 84, 59--65 (1992)
12. Follett, R.H., Follett, R.F.: Use of a Chlorophyll Meter to Evaluate the Nitrogen Status of Dryland Winter Wheat. *Communications in Soil Science and Plant Analysis* 23, 687--697 (1992)
13. Shaahan, M.M., El-Sayed, A.A., Abou El-Nour, E.A.A.: Predicting Nitrogen, Magnesium and Iron Nutritional Status in some Perennial Crops Using a Portable Chlorophyll Meter. *Scientia Horticulturae* 82, 339--348 (1999)
14. Chang, S.X., Robison, D.J.: Nondestructive and Rapid Estimation of Hardwood Foliar Nitrogen Status Using the SPAD-502 Chlorophyll Meter. *Forest Ecology and Management* 181, 331--338 (2003)
15. Li, H.Y., Ren, Q.P., Sun S., Ma, H.Y.: Study on Relation between SPAD Value and Chlorophyll Contents in 10 Kinds Horticulture Woody Plants (in Chinese). *Forestry Science and Technology* 34, 68--70 (2009)
16. Su, Y.S., Guo, H.C., Yang, X.L.: Study on Correlations between SPAD Readings and Chlorophyll Content in Leaves of Sweet Potato, *Dioscorea* and *Konjaku* (in Chinese). *Southwest China Journal of Agricultural Science* 22, 64--66 (2009)

17. Lichtenthaler, H.K., Wellbuen, A.R.: Determinations of Total Carotenoids and Chlorophyll a and b Leaf Extracts in Different Solvents. *Biochemical Society Transactions* 11, 591--592 (1983)
18. Schepers, J.S., Blackmer, T.M., Wilhelm, W., Resende, M.: Transmittance and Reflectance Measurements of Corn leaves from Plants with Different Nitrogen and Water Supply. *Journal of Plant Physiology* 148, 523--529 (1996)
19. Bauerle, W.L., Weston, D.J., Bowdena, J.D., Dudley, J.B., Toler, J.E.: Leaf Absorptance of Photosynthetically Active Radiation in Relation to Chlorophyll Meter Estimates Among Woody Plant Species. *Scientia Horticulturae* 101, 169--178 (2004)
20. Fuchigami, L.H., Ding, P.H., Barnes, G.E.: Portable Meter to Measure Chlorophyll, Nitrogen and Water and Methods. Patent, USA (2007)